

# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <p><b>UKAS</b> CALIBRATION</p> <p><b>0209</b></p> <p>Accredited to <b>ISO/IEC 17025:2017</b></p>	<p><b>Tru-Thread Limited</b></p> <p>Issue No: 023 Issue date: 07 May 2021</p>	
	<p>Unit 3, Roman Park Roman Way Coleshill Birmingham B46 1HG</p>	<p>Contact: Mr S Fisher Tel: +44 (0)1675-462193 Fax: +44 (0)1675-462841 E-Mail: <a href="mailto:quality@tru-thread.co.uk">quality@tru-thread.co.uk</a> Website: <a href="http://www.tru-thread.co.uk">www.tru-thread.co.uk</a></p>
<p><b>Calibration performed at the above address only</b></p>		

### DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty ( $k=2$ )	Remarks
<p>RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED</p>			
<b>LENGTH</b>			<b>NOTES</b>
Plain plug gauges (parallel) and rollers	1 to 50 diameter	0.80	Comparison to gauge blocks using a comparator
	50 to 100	1.0	
	100 to 150	1.5	
	150 to 200	2.0	
	200 to 300	3.0	
Plain plug gauges (taper)	3 to 50 diameter	2.0 on diameter	Comparison to gauge blocks and rollers using a length measuring machine
	50 to 100	3.0	
	100 to 200	4.0	
	200 to 300	5.0	
Taper up to 1 in 8 on diameter	3 to 50 diameter	4.0 on diameter	Using a two axis measuring machine
	50 to 100	5.0	
	100 to 200	6.0	
	200 to 300	7.0	
Taper above 1 in 8 and up to 1 in 3 on diameter	3 to 50 diameter	4.0 on diameter	Using a two axis measuring machine
	50 to 100	5.0	
	100 to 200	6.0	
	200 to 300	7.0	
Plain ring gauges (parallel) and setting standards	2 to 50 diameter	1.5	Comparison to master setting rings using a length measuring machine
	50 to 100	1.8	
	100 to 150	2.0	
	150 to 300	3.0	
Plain ring gauges (taper)	2 to 50 diameter	3.0 on diameter	Comparison to gauge blocks and vee end pieces using an internal diameter measuring machine
	50 to 100	4.0	
	100 to 200	5.0	
	200 to 300	8.0	
Taper up to 1 in 8 on diameter	2 to 50 diameter	5.0 on diameter	Comparison to gauge blocks and vee end pieces using an internal diameter measuring machine
	50 to 100	6.0	
	100 to 200	7.0	
	200 to 300	10	
Taper above 1 in 8 and up to 1 in 3 on diameter	2 to 50 diameter	5.0 on diameter	Comparison to gauge blocks and vee end pieces using an internal diameter measuring machine
	50 to 100	6.0	
	100 to 200	7.0	
	200 to 300	10	



0209  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Tru-Thread Limited**  
Issue No: 023 Issue date: 07 May 2021

Calibration performed at main address only

Measured Quantity Instrument or Gauge	Range	Calibration and Measurement Capability (CMC) Expressed as an Expanded Uncertainty (k=2)	Remarks
RANGE IN MILLIMETRES AND UNCERTAINTY IN MICROMETRES UNLESS OTHERWISE STATED			
LENGTH Continued			
Screw plug gauges (parallel) including check and setting plugs See Note 1	1 to 100 diameter 100 to 300 300 to 400	2.5 on pitch diameter 5.0 7.0	By comparison to cylindrical setting standards and thread measuring cylinders using a screw diameter measuring machine  Note 1. Single and multi-start symmetrical and asymmetrical thread forms.
Screw plug gauges (taper) including API working gauges and API profile gauges to Specs. 5B, 7 & 11B See Note 2	2 to 100 diameter 100 to 300 300 to 350	4.0 8.0 10	By comparison to cylindrical setting standards and thread measuring cylinders using a bench micrometer. API gauge compared to master ring Note 2 Single start, symmetrical thread forms only.
Screw ring gauges (parallel) See Notes 1 and 3	1 to 100 diameter 100 to 150 150 to 300 300 to 400	5.0 6.0 8.0 12	comparison to a cylindrical vee groove setting standard using a two axis measuring machine Note 3. Includes use of check plugs for screw rings from 1 mm to 6mm diameter.
Screw ring gauges (taper) including API working gauges to Specs. 5B, 7 & 11B See Note 2	6 to 150 diameter 150 to 300 300 to 350	7.0 10 15	Using a two axis measuring machine. API gauge compared to master plug
Screw pitch	0.2 to 8	1.5	Using a two axis measuring machine
Screw flank angle	0° to 52°	5.0 minutes of arc	Optical projection and a hob & flank machine
Thread pitch profile gauges	1 to 8 pitch	7.0	Optical projection and a two axis measuring machine
Plain gap gauges (parallel)	2 to 50 50 to 100 100 to 200 200 to 300	3.0 5.0 8.0 12	By comparison to gauge blocks
END			



0209  
Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Tru-Thread Limited**  
Issue No: 023 Issue date: 07 May 2021

Calibration performed at main address only

**Appendix - Calibration and Measurement Capabilities**

**Introduction**

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

**Calibration and Measurement Capabilities (CMCs)**

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest uncertainty of measurement that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors. The CIPM-ILAC definition of the CMC is as follows:

A CMC is a calibration and measurement capability available to customers under normal conditions:

- (a) as published in the BIPM key comparison database (KCDB) of the CIPM MRA; or
- (b) as described in the laboratory's scope of accreditation granted by a signatory to the ILAC Arrangement.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The CMC is calculated according to the procedures given in M3003 and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published CMC in certificates issued under its accreditation.

The CMC may be described using various methods in the Schedule of Accreditation:

As a single value that is valid throughout the range.

As an explicit function of the measurand or of a parameter (see below).

As a range of values. The range is stated such that the customer can make a reasonable estimate of the likely uncertainty at any point within the range.

As a matrix or table where the CMCs depend on the values of the measurand and a further quantity.

In graphical form, providing there is sufficient resolution on each axis to obtain at least two significant figures for the CMC.

**Expression of CMCs - symbols and units**

In general, only units of the SI and those units recognised for use with the SI are used to express the values of quantities and of the associated CMCs. Nevertheless, other commonly used units may be used where considered appropriate for the intended audience. For example, the term "ppm" (part per million) is frequently used by manufacturers of test and measurement equipment to specify the performance of their products. Terms like this may be used in Schedules of Accreditation where they are in common use and understood by the users of such equipment, providing their use does not introduce any ambiguity in the capability that is being described.

When the CMC is expressed as an explicit function of the measurand or of a parameter, this often comprises a relative term (e.g., percentage) and an absolute term, i.e. one expressed in the same units as those of the measurand. This form of expression is used to describe the capability that can be achieved over a range of values. Some examples are shown below. It should be noted that these expressions are *not* mathematical formulae but are instead written in a commonly used shorthand for expressing uncertainties - therefore, for purposes of clarity, an indication of how they are to be interpreted is also provided below.

DC voltage, 100 mV to 1 V: 0.0025 % + 5.0  $\mu$ V

Over the range 100 mV to 1 V, the CMC is 0.0025 %·V + 5.0  $\mu$ V, where V is the measured voltage.

Hydraulic pressure, 0.5 MPa to 140 MPa: 0.0036 % + 0.12 ppm/MPa + 4.0 Pa

Over the range 0.5 MPa to 140 MPa, the CMC is 0.0036 %·p + (0.12·10<sup>-6</sup>·p·10<sup>-6</sup>) + 4.0 Pa, where p is the measured pressure in Pa.

It should be noted that the percentage symbol (%) simply represents the number 0.01. In cases where the CMC is stated only as a percentage, this is to be interpreted as meaning percentage of the measured value or indication.

Thus, for example, a CMC of 1.5 % means 1.5 · 0.01 · i, where i is the instrument indication.